


Schedule of Accreditation

issued by

United Kingdom Accreditation Service

21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

 0593 Accredited to ISO/IEC 17025:2005	Lab3 Limited	
	Issue No: 017 Issue date: 09 March 2009	
	1 Dragon Court Crofts End Road St George Bristol BS5 7XX	Contact: Mr K Wright Tel: +44 (0)117 935 4455 Fax: +44 (0)117 935 4941 E-Mail: kevin.wright@lab3.co.uk Website: www.lab3.co.uk
Calibration performed by the Organisations at the locations specified below		

Locations covered by the organisation and their relevant activities

Laboratory locations:

Location details	Activity	Location code
Address 1 Dragon Court Crofts End Road St George Bristol BS5 7XX	Local contact Mr K Wright Tel: +44 (0)117 935 4455 Fax: +44 (0)117 935 4941 Email: kevin.wright@lab3.co.uk Website: www.lab3.co.uk	Temperature P

Site activities performed away from the locations listed above:

Location details	Activity	Location code
Customer's premises, eg, Hospitals, Laboratories, and Manufacturing Plants	Rotational speed (Centrifuges) Temperature (Sterilizers etc, and Thermal Products)	S S



0593
Accredited to
ISO/IEC 17025:2005

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

Lab3 Limited
Issue No: 017 Issue date: 09 March 2009

Calibration performed by the Organisation at the locations specified

DETAIL OF ACCREDITATION

Measured Quantity Instrument or Gauge	Range	Best Measurement Capability Expressed as an Expanded Uncertainty ($k=2$)	Remarks	Location Code
TEMPERATURE			A list of individual approved engineers is held by the laboratory and by UKAS	
Temperature sensors with indicators or recorders	-35 °C to 55 °C 55 °C to 250 °C 250 °C to 350 °C 350 °C to 650 °C	0.02 °C 0.03 °C 0.35 °C 1.5 °C		P
Liquid-in-glass thermometers	-35 °C to 55 °C 55 °C to 250 °C	0.02 °C 0.03 °C	+ 1/10 scale division	P
Temperature block calibrators	-30 °C to 140 °C -30 °C to 50 °C 50 °C to 140 °C 140 °C to 350 °C 350 °C to 650 °C	0.06 °C # 0.13 °C 0.2 °C 0.3 °C 1.5 °C	# this value only applies to blocks with stirred liquids	P
Temperature controlled autoclaves, media preparators, incubators, sterilizers, ovens, environmental chambers, furnaces, refrigerators, freezers, and liquid baths (inclusive of associated indicators, controllers and recorders, all with sensors, within the specified parameters and ranges)				
	-80 °C to -30 °C -30 °C to 140 °C 140 °C to 250 °C 250 °C to 340 °C 340 °C to 640 °C 640 °C to 1040 °C 1040 °C to 1140 °C	0.8 °C 0.5 °C 0.8 °C 0.9 °C 1.0 °C 1.3 °C 2.5 °C	Single and multipoint monitoring probes Time dependent temperature profiling	S
TIME INTERVAL				
Timers	10 s to 1 hr	2 s		P, S
ROTATIONAL SPEED				
Centrifuges	1000 rpm to 20 000 rpm	1.2%		S

END